



Certificate of Dosimetry Service Approval

Reference: ADS1206

By virtue of Article 24 of S.I. No. 125 of 2000 as amended by S.I. 152 of 2012

Amersham Dosimetry Services
GE Healthcare Limited
The Grove Centre
White Lion Road
Amersham
Buckinghamshire HP7 9LL
UK

is approved by the Radiological Protection Institute of Ireland to provide dosimetry services to GE Healthcare employees in Ireland in pursuance of S.I. 125 of 2000.

This approval is granted subject to the condition that the services are provided within the scope of supply set out in schedule 1.

This approval is granted subject to the requirements for approval continuing to be met.

This certificate shall remain in force until the expiry date specified in this certificate or until revoked in writing by the Radiological Protection Institute of Ireland.

Date of Approval: 19 November 2012

Date of Expiry: 19 November 2017

Signed:

Ann McGarry

Chief Executive Officer

Schedule 1

Dosimetry Service: GE Healthcare Limited

Date of Approval: 19 November 2012

| Dosemeter Make and Model | Technology / Type of Dosimetry | Operational Quantity | Radiation Type | Energy Range | Limitations of Use |
|--|--------------------------------|-----------------------|------------------|--|--------------------|
| Harshaw Type 8855 | TLD / Whole body | H _p (10) | External photons | 20 keV – 1.33 MeV | |
| | | | Beta | 70 keV – 2.28 MeV (E _{βmax}) | |
| | | H _p (0.07) | External photons | 20 keV – 1.33 MeV | |
| | | | Beta | 70 keV – 2.28 MeV (E _{βmax}) | |
| Harshaw EXTRAD TLD-700H chipstrates with PVC finger stall holder | TLD / Extremity | H _p (0.07) | External photons | 20 keV – 1.33 MeV | |
| | | | Beta | 70 keV – 2.28 MeV (E _{βmax}) | |