



Certificate of Dosimetry Service Approval

Reference: ADS1205

By virtue of Article 24 of S.I. No. 125 of 2000 as amended by S.I. 152 of 2012

RRPPS, University Hospitals Birmingham
63 Melchett Road
Kings Norton
Birmingham B30 3HP
UK

is approved by the Radiological Protection Institute of Ireland to provide dosimetry services in Ireland in pursuance of S.I. 125 of 2000.

This approval is granted subject to the condition that the services are provided within the scope of supply set out in schedule 1.

This approval is granted subject to the requirements for approval continuing to be met.

This certificate shall remain in force until the expiry date specified in this certificate or until revoked in writing by the Radiological Protection Institute of Ireland.

Date of Approval: 23 June 2017

Date of Expiry: 31 March 2023

Signed: _____

David Pollard
Programme Manager

Schedule 1

Dosimetry Service: RRPPS, University Hospitals Birmingham

Date of Approval: 27 June 2012

Dosemeter Make and Model	Technology / Type of Dosimetry	Operational Quantity	Radiation Type	Energy Range	Limitations of Use
Panasonic UD802A in PVC pouch	TLD / Whole body	H _p (10)	Photon	15 keV – 7 MeV	Not for neutron fields
			Beta	0.7 – 4.5 MeV (E _{βmax})	
		H _p (0.07)	Photon	15 keV – 7 MeV	
			Beta	0.7 – 4.5 MeV (E _{βmax})	
Harshaw LiF-100	TLD / Extremity	H _p (0.07)	Photon	20 keV – >1 MeV	Not for beta
		H _p (3)	Photon	20 keV – >1 MeV	