

## **Certificate of Dosimetry Service Approval**

*Reference: ADS1205*

By virtue of Article 24 of S.I. No. 125 of 2000 as amended by S.I. 152 of 2012

**RRPPS, University Hospitals Birmingham  
63 Melchett Road  
Kings Norton  
Birmingham B30 3HP  
UK**

is approved by the Environmental Protection Agency to provide dosimetry services in Ireland in pursuance of S.I. 125 of 2000.

This approval is granted subject to the condition that the services are provided within the scope of supply set out in schedule 1.

This approval is granted subject to the requirements for approval continuing to be met.

This certificate shall remain in force until the expiry date specified in this certificate or until revoked in writing by the Environmental Protection Agency.

**Date of Approval:** 23 June 2017

**Date of Expiry:** 23 June 2022

**Signed:**



**Micheál Lehane**

**Director, Office of Radiation Protection and Environmental Monitoring**



# Schedule 1

**Dosimetry Service:** RRPPS, University Hospitals Birmingham

**Date of Approval:** 23 June 2017

Dosemeter Make and Model	Technology / Type of Dosimetry	Operational Quantity	Radiation Type	Energy Range	Limitations of Use
Panasonic UD802A in PVC pouch	TLD / Whole body	H <sub>p</sub> (10)	Photon	15 keV – 7 MeV	Not for neutron fields
			Beta	0.7 – 4.5 MeV (E <sub>βmax</sub> )	
		H <sub>p</sub> (0.07)	Photon	15 keV – 7 MeV	
			Beta	0.7 – 4.5 MeV (E <sub>βmax</sub> )	
Harshaw LiF-100	TLD / Extremity	H <sub>p</sub> (0.07)	Photon	20 keV – 1 MeV	Not for beta
		H <sub>p</sub> (3)	Photon	20 keV – 1 MeV	